

## **AUDIOMETRIC TEST BOOTH CERTIFICATION (MEDICAL SURVEILLANCE)**

Command Owning Booth:

Survey Date:

\_\_\_\_\_

\_\_\_\_\_

### **Audiometric Test Booth Data**

Booth Location:

Significant Operating Conditions:

OBA/SLM Data

Microphone Data

Octave Band Filter

Calibrator Data

Manufacturer:

Manufacturer:

Manufacturer:

Manufacturer:

Model#:

Model#:

Model #:

Model #:

Serial#:

Serial#:

Serial#:

Serial #:

Cal Date:

Cal Date:

Cal Date:

Cal Date:

Field Pre-Cal OK? ☐ Yes ☐ No    Field Post-Cal OK? ☐ Yes ☐ No

### **Field Measurements**

**NOTE: Allowable levels found in NEHC Technical Manual TM 6260.51.99-1(May 1999)**

Octave Band

Max SPL for All

Octave Band SPL

Center Frequency

Tests

Inside Booth

(Hz)

(dB)

(dB)

500

27

Exterior dBA:

1000

29

Exterior dBC:

2000

34

4000

39

8000

41

This booth ☐ **Is** ☐ **Is Not** certified for audiometric testing (Check one)

Comments:

Printed Name/Qualification of Certifier

Signature of Certifier

Certifier's Command: \_\_\_\_\_ Date: \_\_\_\_\_